



October 28, 2005

To: Lin Xiang Qian, Vice-President  
Esco Micro Pte. Ltd. (Biotech Equipment Division)  
21 Changi South Street 1, Singapore 486777

Dear Mr. Lin:

This letter is to recognize that the Esco Micro Pte. Ltd. Biological Safety Cabinet model LB2-4B2 is currently under certification process at NSF and has successfully passed all required NSF/ANSI Standard 49 tests, including:

No	Test	Status
1	Pressure maintenance / soap bubble leak	Passed
2	HEPA filter leak	Passed
3	Inflow velocity	Passed
4	Downflow velocity	Passed
5	Noise level	Passed
6	Light intensity	Passed
7	Vibration level	Passed
8	Smoke pattern	Passed
9	Motor / blower performance	Passed
10	Drain spillage trough leakage	Passed
11	Microbiological personnel protection	Passed
12	Microbiological product protection	Passed
13	Microbiological cross-contamination	Passed
14	Microbiological personnel protection envelope	Passed
15	Microbiological product protection envelope	Passed
16	Rear-to-front and side-to-side resistance to distortion	Passed
17	Work surface deflection	Passed
18	Resistance to tipping	Passed
19	Resistance to overturning	Passed
20	Secondary inflow	Completed

Official listing on the NSF web site is anticipated in November 2005. We also would like to confirm that clinical testing is not required for biosafety cabinets sold in the US.

Sincerely,

Steve Williams

Phone: +1 734 769 5357

Fax: +1 734 827 7170

E-mail: williams@nsf.org